ČESKÝ METROLOGICKÝ INSTITUT

EC type-approval certificate

Number: TCM 128/06 - 4428

Revision 1

Issued by:	**Český metrologický institut** **Okružní 31** **638 00 Brno** **Česká republika**
Notified body:	**1383**
In accordance with:	Annex II, article 1. of the Government Order No. 326/2002 Coll. and the Council Directive 90/384/EEC on non-automatic weighing instruments.
Issued to:	**AXIS Sp. z o.o.** **80-125 Gdańsk** **ul. Kartuska 375 B** **Poland**
In respect of:	**non-automatic weighing instrument, single interval** **class (II)** **Type: AG series** **Max = 50g, 100 g, 200 g, 300 g, 500 g, 600 g, 1000 g, 2000 g, 3000 g, 4000 g (depending on model)** **e = 0,01 g / 0,1 g** **e = 10d** **n ≤ 50000** **Temperature range + 18°C / + 33°C**
Manufacturer:	**AXIS Sp. z o.o., Poland**
Valid until:	**11. September 2016**
Document number:	**0115-CS-A002-06**
Description:	Essential characteristics, approved conditions and special conditions, if any, are described in this EC type-approval certificate. This certificate contains 6 pages.

ČESKÝ METROLOGICKÝ INSTITUT NOTIFIKOVANÁ OSOBA AO 250 1383

Brno, 5. January 2007

RNDr. Pavel Klenovský
Notified Body No. 1383