Sworn Polish-English Translator
Marek Niedźwiecki
Office: ul. Ks. J. Popiełuszki 13, 22-100 Chełm, Poland

T: +48 501 478 921
E: office@translationoffice.pl
W: www.translationoffice.pl

CERTIFIED TRANSLATION FROM POLISH INTO ENGLISH
***

[*logo: Central Office of Measures*]

EU TYPE EXAMINATION CERTIFICATE NO PL 17 008

Issued by: CENTRAL OFFICE OF MEASURES [*GŁÓWNY URZĄD MIAR*]
ul. Elektoralna 2, 00-950 Warsaw

Notified Body: 1440

In accordance with: Regulation of Minister of Economic Development of 2 June 2016 on requirements for non-automatic weighing instruments (Journal of Laws, Item 802) implementing Directive 2014/31/UE of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of non-automatic weighing instruments

Issued to manufacturer: AXIS Sp. z o.o., ul. Kartuska 375B, 80-125 Gdańsk. Poland

In respect of: electronic non-automatic weighing instrument of general use or for direct sale to the public

type:	ATA and ATZ	accuracy class: II
Max:	70 g ÷2200 g	
Min:	50 e or 20 e or 5 e	
e:	0.01 g <e<0.1 g	
n:	< 62 000	
d:	e or 0.1 e	
T:	-Max	

Final statement: Non-automatic weighing instrument satisfies the requirements set out in the regulation of Minister of Economic Development of 2 June 2016 on requirements for non-automatic weighing instruments.

Valid until: 20/11/2027

Reference number: BSM-WR.4410.2.2017.PR.6 Number of pages: 10

The principal characteristics, approval conditions and special regulations, if any, are set out in the Annex, which forms an integral part of the certificate.

Weighing Instruments can be prepared to be used for direct sale to the public.

[*Seal with the phrase "Central Office of Measures" on the rim and the coat of arms of the Republic of Poland in the middle*]

Warsaw, 20/11/2017

[*stamp:*]
p.p. the President of the Central Office of Measures
Maciej Dobieszewski
VICE PRESIDENT
[*illegible signature*]

The President of the Central Office of Measures

This EU type examination certificate may not be reproduced other than in full version. Certificate without signature and seal is not valid.

MAREK NIEDŹWIECKI TŁUMACZ PRZYSIĘGŁY JĘZYKA ANGIELSKIEGO NR TP/30/09